

# NMR Troubleshooting

Please only use these troubleshooting steps if you have been directly instructed to do so by NMR Center Staff. Contact us if you need assistance.

## Topspin

- [Topspin Frozen](#)

---

## Bruker 500

- [Restart ELCB](#)
- [Restart Spectrometer](#)
- [Restart Sample Changer](#)

---

## Bruker 400

- [Restart Spectrometer](#)
- [Restart Sample Changer](#)

---

## Varian 600

- [Using variable temperature](#)
- [Restarting spectrometer](#)
- [Turning on the spectrometer after a power outage](#)
- [Restoring cable contact for carbon tuning](#)

From:

<https://pydio.campus.nd.edu/docs/> - **NMR**

Permanent link:

<https://pydio.campus.nd.edu/docs/doku.php?id=troubleshooting:troubleshooting>

Last update: **2024/06/20 20:04**

