NMR Troubleshooting

Please only use these troubleshooting steps if you have been directly instructed to do so by NMR Center Staff. Contact us if you need assistance.

Topspin

- Topspin Frozen

Bruker 500

- Restart ELCB
- Restart Spectrometer
- Restart Sample Changer

Bruker 400

- Restart Spectrometer
- Restart Sample Changer

Varian 600

- Using variable temperature
- Restarting spectrometer
- Turning on the spectrometer after a power outage
- Restoring cable contact for carbon tuning

Codemeter

- Codemeter Troubleshooting

From:
https://pydio.campus.nd.edu/docs/ - **NMR**

Permanent link:
https://pydio.campus.nd.edu/docs/doku.php?id=troubleshooting:troubleshooting

Last update: **2026/02/20 12:31**